



Control Material CS-MP-SC-1

Sr-90 and Cs-137 in Milk Powder

General Information

Intended use Checking the performance of analytical laboratories engaged in the determination of Sr-90 and Cs-137 in food and other biological samples

Produced and certified by:

Institute of Nuclear Chemistry
and Technology, Dorodna 16
03-195 Warsaw, Poland

Description of the material

Milk powder commercially available was homogenized by mixing for 15 hrs. in a plastic drum rotated in three directions, distributed into PE bags in portion of ca. 500 g each and firmly covered. All bags were firmly closed and sterilized by electron beam radiation from linear accelerator LA-13 with a dose of 28 kGy. Homogeneity was examined by measuring of activity concentration of Sr-90 and Cs-137 in randomly selected containers. Assigning of reference values for activity concentration was carried out by interlaboratory comparison in which 13 radiochemical laboratories took part.

Determination of moisture content

In order to express the activity concentration of radionuclides on a dry-weight basis, moisture content should be determined on a separate subsample of 5 g mass by drying in an oven at 80°C for 27 hours.

Assigning of reference values

Reference values were assigned on the basis of the results provided by several radiochemical laboratories. The radionuclides have to be determined by at least two methods in order to assign reference value. A reference value was calculated using the HISTO software delivered by the IAEA, an uncertainty was evaluated according to ISO 13528:2005 [1], ISO GUM [2] and IUPAC harmonized protocol [3].

Reference values for activity concentration of Sr-90 and Cs-137 in Milk Powder CS-MP-SC-1

Radionuclide **reference value + expanded uncertainty (k=2)**

$$X_{\text{ref}} \pm U^*$$

[Bq kg⁻¹ dry mass]

Cs-137 **1.146 ± 0.107**

Sr-90 **0.91 ± 0.03**

* on 1 November 2010 for ¹³⁷Cs and 18 February 2011 for ⁹⁰Sr

The traceability of this Control Material to the SI units is ensured by the use of certified standard solutions by NPL (the UK's National Measurement Institute) for calibration of measurement equipments.

Long-time stability is monitored during storage. The shelf life of the milk powder material has been established to be 31 December 2020.

References

1. ISO 13528:2005 Statistical methods for the use in proficiency testing by interlaboratory comparison, Geneva, 2005
 2. International Organization for Standardization (ISO), Guide to the Expression of Uncertainty in Measurement, ISO, Geneva, 1993 (corrected and reprinted 1995)
 3. M. Thompson, S.L.R. Ellison, R. Wood, The International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry Laboratories (IUPAC Technical Report), Pure Appl. Chem., 78 (2006) 145
-